



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/080,490	02/22/2002	Ray W. Hathaway	SMI0029.US	8480

7590

08/21/2003

Todd T. Taylor
TAYLOR & AUST, P.C.
142 S. Main St.
P.O. Box 560
Avilla, IN 46710

EXAMINER

BAXTER, JESSICA R

ART UNIT

PAPER NUMBER

3731

DATE MAILED: 08/21/2003

Please find below and/or attached an Office communication concerning this application or proceeding.

NK
17

Office Action Summary	Application No. 10/080,490	Applicant(s) HATHAWAY, RAY W.	
	Examiner Jessica R Baxter	Art Unit 3731	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).
- Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 23 May 2003.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-10 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-10 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- 11) ☐ The proposed drawing correction filed on _____ is: a) ☐ approved b) ☐ disapproved by the Examiner.
If approved, corrected drawings are required in reply to this Office action.
- 12) ☐ The oath or declaration is objected to by the Examiner.

Priority under 35 U.S.C. §§ 119 and 120

- 13) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
* See the attached detailed Office action for a list of the certified copies not received.
- 14) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e) (to a provisional application).
a) ☐ The translation of the foreign language provisional application has been received.
- 15) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. §§ 120 and/or 121.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) Paper No(s). _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449) Paper No(s) _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 112

1. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

2. Claims 1-10 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

3. Regarding claims 1 and 10, the limitation, "each said viewing window including at least one convex circumferential segment as viewed from inside said corresponding viewing window," is unclear, since it is not understood where the convex segment is located. It is understood in the art that one looks through a viewing window and not from inside the window.

4. Claim 1 recites the limitation "said corresponding viewing window" in line 6. There is insufficient antecedent basis for this limitation in the claim.

5. Claim 10 recites the limitation "said corresponding viewing window" in line 8. There is insufficient antecedent basis for this limitation in the claim.

Claim Rejections - 35 USC § 102

6. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Art Unit: 3731

7. Claims 1-7 are rejected under 35 U.S.C. 102(b) as being anticipated by U.S. Patent No. 5,690,634 to Muller et al.

Regarding claims 1, 5, 6 and 7, Muller discloses an orthopaedic reamer comprising a shaft (Column 2 lines 18-20) and a head (FIG. 1) coupled with said shaft, said head including a distal face with a plurality of hemispherically shaped cutting teeth (cutting edges 10) and at least one viewing window extending through said head (openings 5), each said viewing window including at least one convex circumferential segment as viewed from inside said corresponding viewing window (FIGS. 2 and 3).

Regarding claim 2, Muller discloses a plurality of viewing windows (openings 5).

Regarding claim 3, Muller discloses that each said viewing window has a predefined cross-sectional configuration, at least two of said cross-sectional configurations being different (FIG. 5 openings 5).

Regarding claim 4, Muller discloses that at least one viewing window has a cross-sectional configuration with a compound curvature (openings 5).

Claim Rejections - 35 USC § 103

8. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

9. Claims 1-10 are rejected under 35 U.S.C. 103(a) as being unpatentable over U.S. Patent No. 6,102,915 to Bresler et al.

Art Unit: 3731

Bresler discloses a shaft (13) coupled to a hemispherical head (2) that has a plurality of cutting teeth (4) and a viewing window (21,22). The cutting teeth have a generally partial hemispherical shape that allow the user to check contact between the head (2) and bone to ensure a good fit. The windows (21, 22) are configured dependent upon the location of the teeth. Bresler describes the method of use in column 5, which includes placing the device against bone, rotating the cutting head (2) and stopping rotation. Bresler discloses the claimed invention except for the shape of the viewing window including at least one convex circumferential segment as viewed from inside said corresponding viewing window. It would have been an obvious matter of design choice to change the shape of the window to include a convex circumferential segment since such a modification would not have changed the function of the device.

Response to Arguments

10. Applicant's arguments with respect to claims 1-10 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

11. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the

Art Unit: 3731

THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jessica R Baxter whose telephone number is 703-305-4069. The examiner can normally be reached on M-F 8:30AM - 5:00PM.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Michael Milano can be reached on 703-308-2496. The fax phone numbers for the organization where this application or proceeding is assigned are 703-305-3590 for regular communications and 703-305-3590 for After Final communications.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the receptionist whose telephone number is 703-308-0858.

Jessica R Baxter
Examiner
Art Unit 3731


JRB

August 19, 2003


MICHAEL J. MILANO
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700